

ARCH Pre-op Checklist

Step	Image
------	-------

1. Set all surgical functions to neutral.

Traction = 0 mm

Rotation = 0 degrees

Extension / Flexion = Neutral

9. Unlock pin on opposite side of carriage.

Shift the red latch under the red extension pin on the side opposite to where you will be operating the ARCH to keep this pin permanently non-engaged. You can now control extension / flexion using just the pin on your side. **Do not pull red pin before firmly holding black handle**.

